



Zenix SEC-MALS Column Manual

Column Information

Utilizing proprietary surface technologies and 3 μm particle size, Zenix SEC phases are made of uniform, hydrophilic, and neutral nanometer thick films chemically bonded on the high purity and enhanced mechanical stability silica. The proprietary surface technologies allow the chemistry of thin film formation to be well controlled, which results in high column-to-column reproducibility. The nature of the chemical bonding and the maximum bonding density of the thin film benefit Zenix SEC phases with high stability. The uniform surface coating enables high efficiency separation. The narrowly dispersed, spherical silica particles of the Zenix packings for SEC-80, SEC-100, SEC-150 and SEC-300 have nominal pore sizes at 80 \AA , 100 \AA , 150 \AA , and 300 \AA , respectively. With a small particle size of 3 μm and specially designed large pore volume (ca. 1.35 mL/g for Zenix SEC-150 and 300, and ca. 1.1 mL/g for Zenix SEC-100), Zenix phases have achieved unprecedented high separation efficiency and resolution. Zenix SEC columns are packed with a proprietary slurry technique to achieve uniform and stable packing bed density for maximum column efficiency.

Zenix SEC phases are designed to ensure highest resolution and maximum recovery for a very broad range of separation applications. These applications cover large biological molecules, such as proteins and nucleic acids; small biological molecules, such as peptides and oligonucleotides; natural polymers, such as polysaccharides; synthetic polymers; biological cells, such as bacteria and virus; and nanomaterials, such as nanoparticles. Typical applications for Zenix SEC columns are separation and detection in aqueous buffer mobile phases.

System cleaning for MALS Applications

MALS is very sensitive to particles and foreign objects entering the detector and the results can be complicated or confusing. Thorough cleaning of the HPLC system and particles-free mobile phase are critical to get reliable MALS results. Following is the recommended procedure to minimize any interference of foreign particles with the MALS data.

Full HPLC system clean once a week. This includes sonicating the in-bottle filter, precolumn filter, and precolumn housing in 50% aqueous ethanol for 10 minutes. After sonication, rinse everything in HPLC grade water. Reinstall the in-bottle filter but keep the precolumn filter off. Run 100% ethanol through the HPLC system for 20 minutes followed by running water for 20 minutes. Reinstall the precolumn filter after both runs are complete. The system is now ready for buffer. Use freshly made buffer and filter through a 0.2 μm filter just before use.

Daily System clean every night. At the end of everyday operation, run water through the system for 20 minutes followed by running 20% aqueous ethanol for 20 minutes. Leave the system stored in 20% ethanol.

Always use a precolumn filter (0.5 μm stainless steel frit) to trap any residual particles. The flow rate should be within the recommended range of the MALS flow cell. Most of the cases the flow should not exceed 1 mL/min.

Column Stability and Performance

Zenix SEC columns use full coverage bonded silica packing, which allows exceptionally high stability. They are compatible with most aqueous buffers, such as ammonium acetate, phosphate, tris, etc. When 150 mM phosphate buffer at pH 7.0 is used as the mobile phase to run Zenix SEC columns, 300 injections or 1 month of usage has negligible deterioration for Zenix SEC columns.

The neutral and hydrophilic Zenix stationary phases have negligible nonspecific interactions with biological molecules, especially proteins. Combined with their high capacity, Zenix SEC columns enable high efficiency and high recovery separations. A typical quality control chromatogram is shown in Figure 1 for a 7.8x300 mm Zenix SEC column.




Figure 1. Elution profiles of a protein mixture by Zenix SEC-150 and Zenix SEC-300 phases.

Column: 3 μm , 7.8x300 mm; Mobile phase: 150 mM Sodium Phosphate Buffer, pH 7.0; Flow rate: 1.0 mL/min; Temperature: Ambient ($\sim 23^\circ\text{C}$); Detection: UV 214nm; Injection: 10 μL
Sample: 1) Thyroglobulin (1.0 mg/mL), 670 kD; 2) BSA dimer, 132 kD; 3) BSA (1.0 mg/mL), 66 kD; 4) Ribonuclease A (1.0 mg/mL), 13.7 kD, and 5) Uracil (0.1 mg/mL), 120D.

Column Characteristics

Silica: Spherical, high purity (<10 ppm metals)

Particle size: 3 μm

Pore sizes for protein separation:

80 \AA , MW range 100 ~ 50,000

100 \AA , MW range 100 ~ 100,000

150 \AA , MW range 500 ~ 150,000

300 \AA , MW range 5,000 ~ 1,250,000

Safety Precaution

The columns are normally operated under moderate pressure. Loose connections will cause leaking of buffers and injected samples, all of which should be considered as hazards. In the case of leaking, proper gloves should be worn while handling the columns. When opening the columns, proper protections should be used to avoid inhalation of the small silica particles.

Column Installation and Operation

The column should always be capped at both ends when it is not in use. When installing the column to the system, first remove the end caps. Unless a user has special purpose to reverse the flow direction, for example, removal of the inlet blockage, follow the flow direction as marked on the column. Column connections are an integral part of the chromatographic process. If ferrules are over tightened, not set properly, or are not specific for the fitting, leakage can occur. Set the ferrules for column installation to the HPLC system as follows:

(a) Place the male nut and ferrule, in order, onto a 1/16" outer diameter piece of tubing. Be certain that the wider end of the ferrule is against the nut.

(b) Press tubing firmly into the column end fitting. Slide the nut and ferrule forward, engage the threads, and finger-tighten the nut.

(c) Repeat this coupling procedure for the other end of the column.

Samples and Mobile Phases

To avoid clogging the column, all samples and solvents should be filtered through 0.45 μm or 0.2 μm filters before use. Zenix SEC columns are compatible with an aqueous mobile phase or a mixture of organic solvent and water, such as methanol or acetonitrile and water. Always degas the mobile phase. A simple way for degassing is to sonicate it for 5 minutes under water pumped vacuum.

Column Care

Shipping Solvent New columns are shipped in 50 mM sodium phosphate buffer containing 0.02% (w/v) sodium azide (NaN_3). During stocking and shipping, the resin bed may become dried out. To activate a new column, it is recommended to flush the column with 10-20 column volumes of the buffer starting at a low flow rate and gradually increasing until desired operating flow rate is met. Flush the column with your mobile phase while gradually increasing the flow rate to your operating flow rate allowing the column to equilibrate the pressure until the baseline is stable at each step. If the column backpressure and baseline fluctuate, this might be due to air bubbles trapped inside the column. Flush the column with a higher flow rate for 2-5 minutes, for example 0.5 mL/min and 1.25 mL/min for 4.6x300 mm and 7.8x300 mm columns, respectively.

pH For optimum performance and operation during the longest lifetime keep pH between 2 and 8.5.

Pressure Even though the columns can operate at a pressure up to 3,500 psi, the normal operating pressure is usually under 2,000 psi. Continuous use at a high pressure may eventually damage the column. Since the pressure is generated by the flow rate, the maximum flow rate is limited by the backpressure. It is expected that the backpressure might gradually increase with its service. A sudden increase in backpressure suggests that the column inlet frit might be blocked. In this case it is recommended that the column be flushed with reverse flow in an appropriate solvent.

Temperature The maximum operating temperature is 80°C temporarily. The optimum operating temperature for the longest lifetime is 10 - 30°C. Continuous use of the column at a higher temperature (>80°C) can damage the column, especially under high pH (>8).

Flow rate Range Normal operating flow rate is 0.1 - 0.4 and 0.1 - 1.25 mL/min for 4.6 mm and 7.8 mm I.D. columns, respectively.

Storage When the column is not in use for an extended time, the column should be stored in water (H_2O) with 0.02% sodium azide or 10% methanol. Each column is shipped with two removable end plugs. To prevent drying of the column bed, seal both ends of the column with the end plugs provided.

Cleaning From time to time, some samples could get adsorbed onto the inlet frit or the packing material. When the adsorption accumulates to a certain level, it is usually indicated by an increase in back pressure and a broader peak. When this occurs, it is time to clean your column. The general procedure for column cleaning is as follows:

1. Disconnect the column from the detector.

2. Clean your column in the reverse flow direction.

3. Run the column at less than 50% of the maximum recommended flow rate. Monitor the backpressure. If you see the pressure is much higher than the normal operating conditions, you need to lower the flow rate or change the washing buffer as the cleaning solutions may be of different viscosities.

4. Typically, 10-15 column volumes of cleaning solution are sufficient. Rinse well with 3-5 column volume of Nanopure water between each solution.

Cleaning solutions Low pH salt solutions help remove basic proteins. Organics are useful when removing hydrophobic proteins. Chaotropic agents help remove strongly adsorbed materials (e.g., via hydrogen bonding). Only use chaotropic agents when neutral salts or organics have not improved resolution. Two cleaning solutions are recommended for general cleaning:

1. Concentrated neutral salt (e.g., 0.5 M Na_2SO_4) at low pH (e.g., pH 3.0)

2. Water soluble organic (MeOH, ACN, EtOH, 10 %-20 %) in aqueous buffer (e.g., 50 mM phosphate, pH 7.0)

If both solutions fail to clean the column, use 6 M Urea (filter before use).

a. 2 cv 6 M urea at 0.5 mL/min (monitor pressure)

b. 3 cv nanopure water at 0.5 mL/min

c. 7 cv mobile phase at 1 mL/min

Column Protection

In addition to filtering the sample and the mobile phase, the best way to protect the separation column is to install a guard column or a pre-column filter in front of it. In most cases a pre-column filter helps to remove the residual particulates that are in the sample, the mobile phase, or leached from the HPLC system, such as pump and injector seals. However, a guard column is highly recommended because it is more effective in trapping highly adsorptive sample components and residual particulates in the sample, the mobile phase or from the HPLC system.